





America

# CERTIFICATE

No. U8 069850 0006 Rev. 00

**Model(s):** Hei-TORQUE 100 Value Base / Precision Base  
Hei-TORQUE 200 Value Base / Precision Base  
Hei-TORQUE 400 Value Base / Precision Base  
Hei-TORQUE Core / Core Base 400rpm  
And OEM-Variants with other colours or imprint

**Tested according to:** UL 61010-1:2012/R2:2016-04  
CAN/CSA C22.2 No. 61010-1:2012/U2:2016-04  
UL 61010-2-051:2015  
CAN/CSA C22.2 No. 61010-2-051:2015

**Production Facility(ies):** 069850

## Parameters:

Models: Hei-TORQUE 100 Value Base / Precision Base  
Rated Voltage: 100-240 Vac  
Rated Frequency: 50/60 Hz  
Rated Power: 90 W  
Protection Class: I

Models: Hei-TORQUE 200 Value Base / Precision Base  
Rated Voltage: 100-240 Vac  
Rated Frequency: 50/60 Hz  
Rated Power: 120 W  
Protection Class: I

Models: Hei-TORQUE 400 Value Base / Precision Base  
Rated Voltage: 100-240 Vac  
Rated Frequency: 50/60 Hz  
Rated Power: 150 W  
Protection Class: I

Models: Hei-TORQUE Core / Core Base 400rpm  
Rated Voltage: 100-240 Vac  
Rated Frequency: 50/60 Hz  
Rated Power: 105 W  
Protection Class: I

Conditions of acceptability:  
The power supply cord is not in delivery scope.  
For the installation of the equipment, only the recommended power supply cords (indicated in the User manual) shall be used.  
The Marked degree of protection to IEC 60529 was not performed by TÜV Süd.

TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD  
ZERTIFIKAT ◆ CERTIFICATE ◆ 認證證書 ◆ CERTIFICADO ◆ CERTIFICAT

A4 / 07.17